

















KPL-U10060 (WHITE)



SPECIAL FEATURES

- · IP68 rated medical keyboard
- 15 cm shorten cable with USB waterproof silicone cap
- · 4 standard keyboard LEDs (Num, Scroll, Caps lock, CLEAN Mode)
- · Can be cleaned with common detergents
- Additional USB extension cable provided
- Integrated clean mode for cleaning

SPECIFICATION

Number of keys: 105 (US-Layout) Cable length: 15cm with extension cable

Keyswitch technology: membrane Operating temperature: 0°C ~ +55°C Switching cycles: approx. 2 mill. (per key) -15°C ~ +60°C Storage temperature:

Protection level: **IP68** Dimensions: 440(L) x 125(W) x 20(T) mm

Housing material: plastic Weight: approx. 452g

Housing colour: white Certification: CE, FCC, RoHS, REACH

Interfaces: **USB** Warranty: 2 years

ORDER INFORMATION

Cat No.	Product description	Colour	Layout
KPL-U10060	CLEANTYPE® Easy Basic	White	US
KPL-U10060-DE	CLEANTYPE® Easy Basic	White	German
KPL-U10060-AR	CLEANTYPE® Easy Basic	White	Arabic
KPL-U10060-TH	CLEANTYPE® Easy Basic	White	Thai

Other layouts, configurations and interfaces on request.

SUPPLEMENT PRODUCTS

Cat. No.	Product description	Colour
MSI-U10030	Waterproof Medical Touch Scroll Mouse	White
MSI-U20030	Waterproof Medical Touch Scroll Mouse	Black

Mittlerer Ring 1 08233 Treu Germany Tel: E-Mail:

+49 37468 660 0

GETT North America LLC

Suite 203M 2323 S. Voss Road, Houston, TX 77057 United States of America nerica +832 649 7822

GETT Asia Ltd.

Room 806-7, Kowloon Plaza No. 485 Castle Peak Road
Kowloon, Hong Kong
Tel: +852 3461 9789



SHORT OPERATING INSTRUCTIONS



Turn the cleanmode on and off by pressing both keys (duration 5 sec).



LED will be indicated while the cleanmode is on.

CLEANING INSTRUCTION

1) USE DISINFECTANT SPRAYS

Thoroughly spray all surface area and between keys. Wipeaway excess moisture prior to use. It is recommend to use the disinfectant sprays in line with the anufacturer's specifications.

2) USE DISINFECTANT WIPES

Use a disinfectant wipes in-line with manufacturer's specification to clean keyboard. Exert slight pressure and wipe the keyboard surface with circular movement.

For further details please refer to below link:

http://www.gett-group.com/content/service-cleaning-instructions